

ulm university universität
uulm

Physical and Economic Perspectives on Quantum Sensor Technology

22.02.2025

Daniel Dietrich

Dr. Raiker Witter
University Ulm

Contents

1 Introduction **3**

2 Physical Perspectives **4**
2.1 Possible Quantum Sensors 4
2.2 Nitrogen-Vacancy Centers in Diamond 7
2.3 Cold Atom Interferometer Gravimeters 11

3 Economic Perspectives **14**
3.1 Possible Applications of Quantum Sensor Technology 14
3.2 Market Overview 15
3.3 International Competition 17

4 Conclusion **18**

References **20**

1 Introduction

Alongside Albert Einstein's theory of relativity, quantum mechanics is now the most established and best understood theory in modern physics. Quantum mechanics allows extremely precise predictions. Examples include the prediction of the Lamb shift ($5 \cdot 10^{-12}$ eV), the quantum Hall effect ($3 \cdot 10^{-14}$ Siemens) and the anomalous magnetic moment of the electron ($2 \cdot 10^{-13}$). In advanced applications of Nuclear Magnetic Resonance (NMR), particularly in fundamental physics and precision measurements, accuracies on the order of 10^{-16} are achievable, allowing for the detection of minuscule magnetic interactions and the precise determination of fundamental constants. Considering these accuracies, the application of quantum systems as sensors seems obvious. Unlike classical sensors, which are ultimately limited by thermal and electronic noise, quantum sensors can approach the fundamental limits imposed by quantum mechanics itself. The exceptional measurement capabilities of quantum systems stem from several key quantum mechanical properties. First, quantum superposition allows a system to exist simultaneously in multiple states, enabling the detection of extremely weak signals through quantum interference effects. Second, quantum entanglement between particles creates correlations stronger than those possible in classical systems, permitting measurements that surpass classical precision limits. Finally, the discrete nature of quantum states provides a natural quantization that can be exploited for precise calibration-free measurements. These quantum advantages have enabled remarkable breakthroughs in measurement precision, such as magnetic field detection with attotesla sensitivity, time measurements with uncertainties below 10^{-18} seconds, and gravitational field sensing at the micro-g level. Unlike other quantum technologies such as quantum computing or quantum communication that still face significant technical hurdles, quantum sensors are already transitioning from laboratory demonstrations to commercial products. Companies in Europe, North America, and Asia are now competing to bring quantum gravimeters, magnetometers, and atomic clocks to the market. These devices are finding immediate applications in geological surveying, medical imaging and navigation, among other fields. The relatively mature state of quantum sensing technology, combined with its clear commercial value, has sparked intense international competition, with several countries launching dedicated quantum sensing initiatives as part of their national quantum strategies.

2 Physical Perspectives

2.1 Possible Quantum Sensors

This chapter deals with some of the most important quantum sensors. This includes sensors that have only just demonstrated proof of work as well as sensors that already penetrated the commercial market. Figure 1 categorizes these sensors into three groups:

1. Proof of Concept: The technology's feasibility has been demonstrated.
2. Portable Prototype: A functional, portable version of the technology has been developed for testing.
3. Field Use: The technology is already used in real-world applications.

Furthermore, the figure shows the experimentally determined measurement accuracy that can currently be achieved by the quantum sensors in certain applications.

Quantum sensor	functionality	1	2	3	accuracy estimate
Atomic clock	radiative electron transitions	■	■	■	10^{-18} s
Magnetometer	solid state spin defect	■	■		10^{-12} T/$\sqrt{\text{Hz}}$
	vapor cell	■	■	■	10^{-14} T/$\sqrt{\text{Hz}}$
Gravimeter	cold atom interferometer	■	■	■	10^{-8} g
Gyroscope	ultra cold atom interferometer	■			10^{-8} rad/s
	thermal atom beam interferometer	■	■		10^{-2} rad/s
	nuclear magnetic resonance in vapor cell	■	■		10^{-3} rad/s
Accelerometer	cold atom interferometer	■			10^{-9} m/s^2
	thermal atom beam interferometer	■	■		10^{-5} m/s^2
Electrometer	rydberg atoms in vapor cell	■	■		10^{-8} V/cm/$\sqrt{\text{Hz}}$
Thermometer	NV center in diamond	■	■		10^{-3} K
	quantum dots	■	■		10^{-1} K

Figure 1: Overview of the most important quantum sensors with classification into the categories 1: Proof of Concept, 2: Portable Prototype, 3: Field Use and the approximate measurement accuracy currently achievable. [4][15][7][16][10][18][11]

Atomic clocks are one of the most precise measuring instruments known today. They measure time by observing the resonance frequency of atoms. The high measurement accuracy is based, among other things, on the fact that the transition states of the atoms used are insensitive to temperature and other environmental influences. The most common type of atomic clock is the cesium atomic clock, which uses the transition between

two hyperfine levels of the cesium-133 atom. When cesium atoms are exposed to microwave radiation at a frequency of 9,192,631,770 Hz, they undergo a transition that can be detected and used to regulate the clock. [12] This frequency is so stable that it defines the international standard for the second. Their unparalleled measuring accuracy and highly developed miniaturization, such as in chip-scale atomic clocks (CSACs), make atomic clocks a versatile and irreplaceable technology today.




Figure 2: This photograph shows a vapor cell of a chip scale atomic clock (CSAC) compared to the size of a coffee bean. Source: [1]

Atomic vapor-cell magnetometers, also known as optically pumped magnetometers, use the interaction between light and atoms in a vapor cell to measure magnetic fields. When exposed to a magnetic field, the atoms' spins precess at a frequency proportional to the field strength. By detecting this precession frequency, the magnetometer can determine the magnetic field with high accuracy. Recent advances have miniaturized these devices, making them more practical for various applications, including biomedical imaging and geophysical surveys. Magnetoencephalography, typically conducted by superconducting magnetometers, can now be performed using atomic vapor cell magnetometers, which require orders of magnitude less power and volume because of the absence of cryogenic cooling [14]. Quantum magnetometers, capable of resolving magnetic fields of approximately 10^{-12} Tesla within one second of measurement, are available today for a few thousand dollars. These devices are currently replacing less-sensitive classical magnetometers like fluxgate magnetometers, which are offered at similar prices and are being utilized in communication, navigation and geological surveys [14].

The defects in the diamond lattice of a NV center in diamond magnetometer exhibit spin states that are highly sensitive to magnetic fields. By optically initializing and reading out the spin states, NV center in diamond magnetometers can measure magnetic fields with high spatial resolution and sensitivity. This makes them ideal for applications requiring nanoscale magnetic imaging, such as material science and biological research.

Cold-atom interferometer gravimeters measure local gravitational acceleration by measuring the gravitational influence on a matter wave interferometer. The extremely accurate measurements of the gravitational field allow the detection of changes in underground mass distribution. Today, the need for calibration in this technology is largely eliminated and quantum gravimeters are used for oil, gas, water, and mineral prospecting, monitoring volcanic activity and geophysical studies. Due to the effects of the general theory of relativity, atomic clocks can also be used to measure gravitational fields.

Quantum gyroscopes measure the Sagnac effect to determine rotational speed. Ultracold atom interferometer gyroscopes and thermal atom beam interferometer gyroscopes measure the Sagnac effect by detecting the resulting phase shift in the interferometer. Nuclear magnetic resonance in vapor cell gyroscopes on the other hand measure the Sagnac effect by its influence on the precision frequencies of the spins.

Cold atom interferometer accelerometers and thermal atom beam interferometer accelerometers can be viewed as an extension to the quantum gravimeter. The underlying principle of measuring the influence of acceleration/gravitation on a matter wave interferometer is the same. But while quantum gravimeters are made to detect small changes in gravitational acceleration with high precision, quantum accelerometers measure a wide range of accelerations and are expected to be robust and able to work in dynamic environments in the future.

Quantum electrometers use atoms in a vapor cell, typically rubidium or cesium, that are excited to Rydberg states using laser light. These Rydberg states have large electric dipole moments, making them highly sensitive to electric fields. The vapor cell is subjected to a probe laser and a coupling laser. The interaction between these lasers and the Rydberg atoms creates an electromagnetically induced transparency (EIT) condition, where the medium becomes transparent to the probe laser at specific frequencies. When an external electric field is applied, it shifts the energy levels of the Rydberg atoms, altering the EIT condition. This change in transparency is detected by measuring the transmission of the probe laser through the vapor cell. Both the magnitude and the direction of the electric field can be determined with the quantum electrometer. Quantum electrometers can be used for high-precision radio frequency (RF) detection. They offer comprehensive spectral coverage and a wide dynamic range that exceeds the capabilities of conventional receiver antennas [14].

NV centers in diamonds can be used as quantum thermometers. This is because the shifts in the energy levels of NV centers are highly temperature-dependent. Quantum dot thermometers use semiconductor nanocrystals (quantum dots) that have size-dependent optical and electronic properties due to quantum confinement effects. The temperature is determined by monitoring the temperature-dependent shifts in the photoluminescence spectra of quantum dots. Quantum thermometers are able to monitor sub-kelvin variations over a wide temperature range.

2.2 Nitrogen-Vacancy Centers in Diamond

Quantum sensing usually employs a negatively charged Nitrogen vacancy defect structure (NV center) in diamond, therefore the negatively charged NV center is explicitly discussed in this chapter. NV centers in diamond have emerged as a groundbreaking tool for quantum sensing. Especially due to their high sensitivity to magnetic fields, electric fields and temperature. Alternative defects in silicon carbide, yttrium orthovanadate, and rare-earth-doped crystals have also been investigated. However, diamond remains superior to this day. Due to its properties, the diamond lattice is ideal for quantum experiments. Diamond has a broadband optical transparency and extremely high thermal conductivity. It withstands high electric fields and, although it is an electrical insulator, it behaves like a semiconductor when doped. In the negatively charged NV center, one carbon atom is replaced by a nitrogen atom and an adjacent carbon atom is replaced by a vacancy that confines two electrons.




Figure 3: Simplified depiction of a NV center in the diamond lattice. The carbon atoms are green, the nitrogen atom is red and the vacancy is marked with a “V”.

For the production of NV centers, high purity diamond is used. It is usually synthesized using chemical vapor deposition (CVD). Nitrogen atoms are integrated either during CVD growth or by subsequent ion implantation, with the latter method allowing more precise control over the position of the defects. The required lattice vacancies are created by irradiation with high-energy particles, typically electrons or protons. In a subsequent annealing process at temperatures between 600 and 700 ℃, these vacancies are mobilized and can recombine with the implanted nitrogen atoms to form stable NV centers [17].

The electronic structure of the center, shown in figure 4, consists of a ground triplet state ${}^3A_2$, an excited triplet state ${}^3E$ and several dark singlet states. The ground triplet state exhibits a remarkably stable zero-field splitting of $2.87\,\mathrm{GHz}$ between the $m_s = 0$ and $m_s = \pm 1$ spin sub-levels, showing only weak dependence on both temperature and crystal strain. In contrast, the fine structure of the excited triplet state ${}^3E$ is highly dependent on

temperature and crystal strain. But at room temperature, due to the dynamic Jahn-Teller effect, it displays a strain-independent zero-field splitting of 1.42 GHz [6].




Figure 4: The electronic orbital structure (left) and fine structure (right) of a NV- center in diamond at room temperature. The observed optical zero phonon line (1.945 eV) and infrared zero phonon line (1.190 eV) transitions are depicted as arrows in the orbital structure.

The ground state Hamiltonian of the negatively charged NV center in diamond can be written as:

$$\hat{H} = D(S_z^2 - \frac{1}{3}S(S+1)) + E(S_x^2 - S_y^2) + \gamma\vec{B}\vec{S} + \sum_i A_i S I_i \tag{1}$$

The first term with the parameter $D = 2.87$ GHz represents the zero-field splitting due to spin-spin interactions, where S is the total electron spin and S_z is its projection along the NV symmetry axis. For the negatively charged NV center, the total electron spin is $S = 1$. The second term describes the strain-induced splitting, where S_x and S_y are the spin operators and E is the strain-induced splitting parameter. The third term represents the Zeeman interaction with an external magnetic field $\vec{B}$, where γ is the electron gyromagnetic ratio. The final term describes the hyperfine interaction between the electron spin S and the nuclear spins I_i of the nitrogen atom and the nearby carbon atoms. A_i is the hyperfine coupling tensor for the i-th nuclear spin. At low magnetic fields, zero-field splitting

is the dominant interaction [6]. In this representation of the Hamiltonian, anisotropic terms and higher orders of the electron-electron spin-spin interaction are neglected. In addition, higher order magnetic fields, phonon couplings, anisotropic strain effects or distortions in the crystal and higher order hyperfine interactions are neglected. These include quadrupolar nuclear spin interactions, nucleus-nucleus interactions, nonlinear effects due to strong magnetic fields and interactions between electron spin and distant nuclei across the crystal lattice (superhyperfine interactions). These additional terms can usually be neglected because their effects are typically much smaller than the dominant interactions included in the Hamiltonian. For example, quadrupole interactions and dipole-dipole couplings are weaker by orders of magnitude compared to zero-field splitting or Zeeman interactions. Similarly, higher-order terms and external field effects (electric or magnetic) are negligible under standard experimental conditions, where these perturbations are small or well controlled.

For the use of this quantum system as a sensor, the diamond pumped with NV centers is excited with a green laser (532nm). This induces a spin-conserving transition of the NV center from the ground state $^3A_2$ to the excited triplet state ^3E. From this excited state, there are two competing relaxation paths: The primary, radiative transition leads directly back to the $^3A_2$ ground state with the emission of red fluorescence, characterized by a zero phonon line at $637\,\text{nm}$ and adjacent phonon sidebands up to about $800\,\text{nm}$. This spin-conserving process dominates in particular for the $m_s = 0$ state. In parallel, there is a non-radiative relaxation path via the singlet system, which is preferentially followed from the $m_s = \pm 1$ states. This path leads via an intersystem crossing cascade from the ^3E to the $^1A_1$ singlet state and further via the ^1E state back to the $^3A_2$ ground state, whereby this process preferentially ends in the $m_s = 0$ state.

These different transition probabilities enable the optical readout of the spin state: The $m_s = 0$ state shows a high fluorescence intensity due to the dominant radiative transition, while the $m_s = \pm 1$ states show reduced fluorescence due to the enhanced non-radiative relaxation path. In optically detected magnetic resonance (ODMR) spectroscopy, this spin-dependent fluorescence is utilized by applying a microwave field of variable frequency during continuous optical excitation. If the microwave frequency coincides with the transition energies between $m_s = 0$ and $m_s = \pm 1$, the ground state population shifts to the darker $m_s = \pm 1$ states, which leads to characteristic dips in the fluorescence spectrum. An external magnetic field $\vec{B}$ lifts the degeneracy of the $m_s = \pm 1$ states due to the Zeeman effect, causing the resonance frequencies to shift.

For the weak magnetic field regime, the electron spin resonance (ESR) frequencies $\nu_{\pm}$ are given by:

$$\nu_{\pm}(B) = D \pm \sqrt{\left(\frac{g\mu_B}{h}B\right)^2 + E^2} \tag{2}$$

μ_B is the Bohr magneton and g is the Landé g-factor for the electron. B is the strength of the external magnetic field, D is the zero-field splitting and E is the strain-induced splitting parameter. The zero-field splitting is strongly influenced by the temperature and can be approximated by the following equation:

$$D(T) = D_0 + \alpha T + \beta T^2 \tag{3}$$

Here is $D_0 \approx 2.87\,\text{GHz}$, $\alpha \approx -74\,\text{kHz/K}$ and $\beta \approx -0,35\,\text{kHz/K}$. The figure 5 shows the ODMR spectrum of a single NV center in diamond.




Figure 5: ODMR spectrum of a single NV center without external magnetic field and with external magnetic field. A fit curve (black) was applied to the data points (red). Source: [9]

Figure 6 shows the basic structure of an NV center quantum sensor.




Figure 6: experimental setup of an NV center quantum sensor

The experimental apparatus combines optical and microwave control elements: a confocal microscope system with typical laser excitation at 532 nm enables optical initialization and readout of the NV spin state, while microwave delivery through proximal antennas allows coherent manipulation of the spin states. Fluorescence detection is achieved using single-photon counting modules in the 600-800 nm range. The diamond sample is typically mounted on precision positioning stages within a controlled environment, where magnetic fields, temperature, or other target parameters can be systematically varied. This configuration enables the NV center to function as a quantum sensor through measurement of Zeeman shifts, strain effects, or temperature-dependent zero-field splitting, with demonstrated sensitivities reaching the nanoscale regime across multiple physical modalities.

2.3 Cold Atom Interferometer Gravimeters

Quantum gravimeters harness the wave-like nature of matter to perform ultra-precise measurements of gravitational acceleration. Unlike classical gravimeters, which rely on mechanical or optical measurements of test masses, quantum gravimeters exploit the quantum mechanical behavior of atoms to achieve superior precision and long-term stability.

The core principle of quantum gravimeters relies on matter-wave interferometry, typically implementing a Mach-Zehnder-type interferometer configuration commonly using rubidium atoms.




Figure 7: simplified depiction of the path of the matter wave through space-time in a cold atom interferometer gravimeter. Excited atoms are depicted red, atoms in the ground state blue.[13]

The interferometer begins with the preparation of atoms in a well-defined quantum state, typically utilizing the hyperfine ground states of alkali atoms. For example, in 87Rb, the $|F = 1, m_F = 0\rangle$ and $|F = 2, m_F = 0\rangle$ states serve as the fundamental two-level system. The coherent manipulation of these states is achieved through stimulated Raman transitions, employing two counter-propagating laser beams whose frequency difference matches the hyperfine splitting plus the Doppler shift and recoil shift. The Raman transition process can be described by the interaction Hamiltonian:

$$\hat{H} = \frac{\hbar\Omega_{eff}}{2}\left(|e,p\rangle\langle,p+\hbar k|e^{i\phi} + h.c.\right) \tag{4}$$

where Ω_{eff} is the effective Rabi frequency, $|g\rangle$ and $|e\rangle$ represent the ground and excited states respectively, k is the effective wavevector, and ϕ is the laser phase difference.

The interferometer sequence consists of three pulses: Initially, a $\frac{\pi}{2}$-pulse (beam splitter) with a typical duration of $\tau \approx 10 - 50\,\mu$s creates a quantum superposition of states with different momenta, effectively splitting the atomic wavefunction into two paths[13]. The resulting state can be described as $\frac{1}{\sqrt{2}}(|g,p\rangle + |e,p+\hbar k\rangle)$. After a free evolution time T, a π-pulse (mirror) with the duration 2τ reverses the relative motion of the two paths and inverts the population between $|g,p\rangle$ and $|e,p+\hbar k\rangle$. After another time T, a final $\frac{\pi}{2}$-pulse

recombines the paths and converts the accumulated phase difference into a population difference. In a uniform gravitational field, the gravitational phase shift is given by:

$$\Delta\phi_g = g \cdot k \cdot T^2 \tag{5}$$

The quadratic dependence on the interrogation time T provides the high sensitivity of atomic gravimeters. The interferometer output is determined by measuring the population distribution between the two states after the final $\frac{\pi}{2}$-pulse. The transition probability follows:

$$P = |\langle e|\Psi\rangle|^2 = \frac{(1 - \cos(\Delta\phi_{total}))}{2} \tag{6}$$

The population difference is typically measured through state-selective fluorescence detection, where atoms are illuminated with a resonant probe beam and scattered photons are collected by photodetectors. The normalized population difference provides a direct measure of gravitational acceleration:

$$N = \frac{N_e - N_g}{N_e + N_g} = C \cdot \cos\left(k \cdot g \cdot T^2 + \phi_0\right) \tag{7}$$

C is the contrast of the interference fringes and ϕ_0 is an additional phase offset.

Various techniques are employed to enhance sensitivity: Large Momentum Transfer (LMT) beam splitters increase the effective momentum separation through multiple photon recoils, enhancing the phase shift by a factor of up to 100. Extended free-fall times through atomic fountains or drop towers exploit the quadratic sensitivity improvement with T. Delta-kick collimation reduces atomic ensemble expansion through pulsed magnetic fields or optical potentials, enabling longer interrogation times.

The accuracy of matter-wave interferometers is limited by several systematic effects: AC Stark shifts from Raman lasers (compensation through intensity ratios), magnetic field gradients (second-order Zeeman effect, requiring magnetic shielding), wavefront aberrations (phase errors from non-ideal optical wavefronts), and the Coriolis effect due to Earth's rotation (compensation through counter-propagating atomic trajectories). The implementation of matter-wave interferometry in quantum gravimeters represents a remarkable achievement in precision measurement, combining quantum mechanics, atomic physics, and optical manipulation techniques to achieve unprecedented sensitivity in gravitational acceleration measurements.

3 Economic Perspectives

3.1 Possible Applications of Quantum Sensor Technology

Quantum sensing technologies are revolutionizing measurement capabilities across industries. These advances create unprecedented opportunities in sectors ranging from mineral exploration to healthcare, with significant economic implications. The following figure 8 shows some possible applications of this technology [5]:




Figure 8: present and future applications of quantum sensor technology [5][4][7].

In the financial sector, atomic clocks have become indispensable for high-frequency trading systems, providing nanosecond-precise timestamps for transaction synchronization and regulatory compliance. These same clocks are crucial for telecommunications, particularly in 5G infrastructure, where they maintain network synchronization and data integrity. Global Navigation Satellite Systems like GPS rely heavily on atomic clock precision. Looking ahead, miniaturized atomic clocks could transform autonomous vehicle navigation, especially in GPS-denied environments, while also enhancing financial market surveillance and fraud detection capabilities. Quantum magnetometers are making waves in mineral exploration, medical imaging, and materials science. Their ability to detect magnetic anomalies aids in resource identification, while their application in magnetoencephalography enables detailed brain activity mapping. Future applications could revolutionize non-invasive medical diagnostics, underground infrastructure mapping, and space weather monitoring for protecting power grids from solar storms. The field of quantum gravimetry is particularly impactful in geological surveys, civil engineering, and resource

exploration. These devices excel at detecting underground structures and assessing stability. Their potential extends to critical environmental applications, such as monitoring ice mass changes for climate research and mapping groundwater resources. The technology could also provide early warning systems for structural instabilities in critical infrastructure. Quantum gyroscopes are enhancing navigation systems and platform stabilization, particularly in aerospace and marine applications. Their future applications could revolutionize indoor navigation, autonomous vehicle systems, and space exploration through improved spacecraft orientation control. Similarly, quantum accelerometers are advancing inertial navigation, seismic monitoring, and structural health assessment. Their potential extends to personal navigation systems, industrial automation, and transportation safety through advanced collision avoidance systems. In the realm of electrical measurements, quantum electrometers are proving valuable for materials characterization and semiconductor manufacturing. Their future applications could enhance brain-computer interfaces and power grid monitoring. Meanwhile, quantum thermometers are pushing the boundaries of precision temperature measurement in manufacturing, medical diagnostics, and scientific research. Their potential applications include early cancer detection through precise temperature mapping and improved industrial process control [5].

3.2 Market Overview




Figure 9: Market share diagram of the quantum sensing market with atomic clocks, magnetometers, gyroscopes and accelerometers, gravimeters, electrometers and thermometers. The total volume amounts to $478 million.

The quantum sensing market is experiencing significant growth, driven by technological advancements and increasing demand across multiple sectors. As of 2025, the global quantum sensing market is valued at approximately $478 million, with projections indicating a compound annual growth rate (CAGR) of 15.2% through 2030, potentially reaching $1.2 billion.

The atomic clock segment currently dominates the quantum sensing market, accounting for roughly 35% of the total market share. This dominance is primarily attributed to their essential role in telecommunications and financial services. The chip-scale atomic clock market alone is expected to reach $325 million by 2028, growing at a CAGR of 8.7%. Major telecommunications companies are investing heavily in atomic clock technology, with individual units ranging from $50,000 to $500,000 depending on precision requirements. Quantum magnetometers represent the second-largest market segment, currently valued at approximately $156 million. The medical applications of magnetometers, particularly in magnetoencephalography (MEG), are driving significant growth. A single MEG system can cost between $2 million and $3 million, with the market for medical quantum magnetometers expected to grow at a CAGR of 16.8% through 2030 [2].

The quantum gravimeter market, though smaller at present ($89 million), shows remarkable growth potential. The oil and gas industry's increasing adoption of quantum gravimetry for resource exploration is expected to drive the market to $275 million by 2030. Individual quantum gravimeter units currently range from $100,000 to $400,000, with prices expected to decrease significantly as manufacturing scales up. Quantum gyroscopes and accelerometers together constitute a $120 million market segment, with significant growth potential in autonomous navigation systems. The inertial navigation system market, which heavily relies on these technologies, is projected to reach $1.8 billion by 2030, with quantum sensors capturing an increasing share of this market. The quantum electrometer market, currently valued at $45 million, is experiencing rapid growth due to increasing demand in semiconductor manufacturing. Market projections suggest a CAGR of 18.5% through 2030, potentially reaching $180 million. This growth is partly driven by the expansion of the semiconductor industry, which is expected to reach $1 trillion by 2030. Quantum thermometry represents the smallest but fastest-growing segment, with a current market value of $35 million and projected CAGR of 22.3% through 2030. The medical applications of quantum thermometers, particularly in cancer detection, are expected to drive significant growth in this segment [2].

Regional analysis indicates that North America currently holds the largest market share (38%), followed by Europe (29%) and Asia-Pacific (24%). However, the Asia-Pacific region is expected to show the highest growth rate, with a projected CAGR of 17.8% through 2030, driven by increasing investments in quantum technology by countries like China, Japan, and South Korea. Overall, investment in quantum sensing research and development has seen substantial growth, with government funding worldwide exceeding $2.5 billion in 2023. Private sector investment has also increased significantly, with venture capital funding in quantum sensing startups reaching $850 million in 2023, a 45% increase from the previous year. Market barriers include high initial costs and technical complexity. The average cost of implementing quantum sensing solutions ranges from $500,000 to $5 million, depending on the application and precision requirements. However, these costs are expected to decrease by 30-40% over the next five years as manufacturing processes improve and economies of scale are realized.

The competitive landscape is characterized by a mix of established technology companies and specialized quantum technology firms. Major players include Honeywell (US), Microsemi (US), Muquans (France), AOSense (US), and M Squared (UK), collectively holding approximately 65% of the market share. Recent years have seen increased merger and acquisition activity, with deals totaling $1.2 billion in 2023 alone. Emerging applications in environmental monitoring and climate change research are expected to create new market opportunities. The market for quantum sensors in environmental monitoring is projected to reach $150 million by 2030, driven by increasing demand for precise measurements of greenhouse gases and environmental parameters. End-user industry analysis shows that defense and military applications currently account for 28% of the market, followed by healthcare (22%), telecommunications (18%), and natural resource exploration (15%). However, commercial applications are expected to show the highest growth rate, with a projected CAGR of 19.2% through 2030. The impact of technological advancement on market growth is significant. Manufacturing costs for quantum sensors are decreasing at an average rate of 12% annually, while performance metrics continue to improve by 15-20% per year. This trend is expected to accelerate market adoption across various industries.

3.3 International Competition

International competition in quantum sensing technology has intensified significantly over the past decade, with major global powers investing heavily in research, development, and commercialization. This technological race has become a crucial element of national security and economic strategy, particularly among the United States, China, and the European Union.

Current investment patterns reveal the strategic importance placed on quantum sensing technology. The United States commits approximately $1.2 billion annually to quantum research, with about $300 million specifically allocated to sensing technologies. Private sector engagement is particularly strong in the US, with venture capital funding reaching $850 million in 2023. China has reportedly invested even more aggressively, announcing a $15 billion investment over five years in quantum technologies, although this figure includes broader quantum applications beyond sensing, and the actual allocation specifics remain unclear. The European Union has taken a more measured approach through its Quantum Flagship program, committing $1 billion over ten years, supplemented by approximately $5 billion from individual member states.

Patent filings provide another metric for assessing technological progress, though these numbers should be interpreted cautiously. While China leads with over 2000 quantum sensing-related patents, questions persist about their novelty and quality. The United States has filed approximately 1200 patents in the past five years, demonstrating strong innovation, particularly in commercial applications. European entities have filed around 800 patents, often focusing on fundamental research and precision engineering applications.

The technological landscape shows distinct regional strengths. The United States leads in atomic clock technology and defense applications, supported by strong private-public partnerships and a robust venture capital ecosystem. American companies dominate in commercialization efforts, particularly in integrating quantum sensors into existing technologies. However, the US faces challenges in coordinating research efforts and addressing a growing skills shortage in quantum engineering. China's centralized approach has yielded rapid advancement in quantum network infrastructure and manufacturing capabilities. Their claimed workforce of 50,000 quantum researchers (though classification criteria remain unclear) demonstrates significant human capital investment. However, China continues to depend on foreign technology in key areas and faces challenges in quality control and international collaboration due to security concerns. European excellence lies in fundamental research and precision measurement applications, particularly in quantum gravimetry and magnetometry. The strong international collaboration networks and established research infrastructure in the EU provide a solid foundation for advancement. However, the region struggles with fragmented funding across member states and slower commercialization compared to its competitors.

Looking forward, success in this technological race will likely depend on addressing several critical factors. The United States could strengthen its position by establishing more coordinated research efforts and securing supply chains for critical components. For China, focusing on indigenous innovation capabilities and improving international research integration would be beneficial, while maintaining transparency to build trust in the global scientific community. The European Union could enhance its competitive position by accelerating commercialization efforts and developing stronger retention programs for talented researchers.

4 Conclusion

Quantum technologies utilize the laws of quantum mechanics to achieve things that are often difficult or impossible to imagine in classical physics. Quantum sensors are an interesting application because they enable high-precision measurements and are often particularly stable or even drift-free.

The global quantum sensor market is flourishing and will experience strong growth over the next few years. From 2025 to 2030, it is expected to increase from $478 million to $1.2 billion. Atomic clocks currently dominate the market with a market share of 35%, followed by quantum magnetometers. Different focal points are emerging in international competition: The USA, which invests $1.2 billion annually in quantum sensing, is leading the commercialization of atomic clock technology. China is focusing heavily on quantum network infrastructure with an investment of $15 billion over five years, while the EU stands out with its billion-dollar investment in basic research and precision measurement.

Artificial intelligence (AI) may also play a transformative role in the future development of quantum sensors. Especially in the optimization of sensor designs, AI could model com-

plex quantum systems and identify the optimal parameters for specific applications without having to fully calculate the system. This would significantly shorten the traditionally time-consuming process of experimental optimization. In the production of quantum sensors, AI-controlled systems could revolutionize quality control by detecting defects more precisely. Production parameters could also be adjusted in real time to increase the yield of high-quality sensors. In addition, AI could also play a role in the calibration or even operation of quantum sensors. A promising approach could also be for AI to identify new applications for quantum sensors. In the distant future, quantum computers could also become interesting for describing quantum systems with mathematical precision and suggesting the best systems for quantum sensors. In this respect, the development of quantum technologies may show promising self-enhancing properties.

In contrast to other quantum technologies, quantum sensors are already a mature technology in many areas with immediate commercial applications. The further development in the industrialization of quantum sensor technology will accelerate due to already foreseeable application possibilities, which will be driven by the expansion of a global value chain for quantum technologies. This also makes quantum sensors an important long-term milestone in the commercialization of various quantum technologies, including quantum computers.

References

[1] *Chip Scale Atomic Clock*. Wikipedia, 2023

[2] *Analyse der Marktgröße und des Anteils von Quantensensoren - Branchenforschungsbericht - Wachstumstrends*. Mordor Intelligence™ Industry Reports, 2024

[3] *Marktgröße, Anteil und Branchenanalyse für Quantensensoren, nach Typ, Branche und Region, 2024–2032*. Fortune Business Insights, 2025

[4] ALSID, S. T. ; SCHLOSS, J. M. ; STEINECKER, M. H.: *Solid State Microwave Magentometer with Picotesla Level Sensitivity*. American Physical Society, 2023

[5] BORMUTH, Y. ; SOLLER, H. ; GSCHWENDTER, M. : *Quantum Sensing Can Already Make a Difference.But Where?* Journal of Innovation Management, 2024

[6] DOHERTY, M. W. ; DOLDE, F. ; JELEZKO, F. : *Theory of the ground state spin of the NV- center in diamond: I. Fine structure, hyperfine structure, and interactions with electric, magnetic and strain fields*. arXiv, 2011

[7] GAUGUET, A. ; CANUEL, B. ; LEVEQUE, T. : *Characterization and limits of a cold-atom Sagnac interferometer*. arXiv, 2018

[8] GILLOT, P. ; CHENG, B. ; IMANALIEV, A. : *Cold Atom Gravimeters*. SYRTE Observatoire de Paris, 2021

[9] HONG, S. ; GRINOLDS, M. S. ; PHAM, L. M.: *Magnetometry with single NV centers in diamond*. National University of Singapur, 2025

[10] LEE, S. ; LEE, D. Y. ; SHIM, K. M.: *Performance comparison between 87Rb-natural Xe-N2 and 87Rb-129Xe-131Xe-N2 atom spin gyroscopes*. Korean Physical Society, 2021

[11] LIU, C. F. ; LEONG, W. H. ; XIA, K. : *Ultra-sensitive hybrid diamond nanothermometer*. arXiv, 2019

[12] LUDLOW, A. D. ; BOYD, M. M. ; YE, J. : *Optical Atomic Clocks*. arXiv, 2015

[13] MENORET, V. ; VERMEULEN, P. ; LEMOIGNE, N. : *Gravity measurements below* 10^9 *g with a transportable absolute quantum gravimeter*. scientific reports, 2018

[14] OH, E. ; GREGOIRE, M. D. ; BLACK, A. T.: *A Perspective on Quantum Sensors from Basic Research to Commercial Applications*. arXiv, 2024

[15] RAGHAVAN, H. ; TAYLER, M. C. D. ; MOULOUDAKIS, K. : *Functionalized millimeter-scale vapor cells for alkali-metal spectroscopy and magnetometry*. American Physical Society, 2024

[16] SATO, T. ; NISHIMURA, N. ; KAKU, N. : *Closed-loop measurements in an atom interferometer gyroscope with velocity-dependent phase-dispersion compensation*. arXiv, 2025

[17] TOURAL, J. L. S. ; MARZOA, V. ; BERNARDO-GAVITO, R. : *Hands-On Quantum*

Sensing with NV Centers in Diamonds. Multidisciplinary Digital Publishing Institute, 2023

[18] TRAVAGNIN, M. : *Cold atom interferometry sensory: physics and technologies*. Publications Office of the European Union in Luxembourg, 2020